

References

Al-Mezaine HS, Al-Amro SA, Kangave D, Sadaawy A, Wehaib TA, Al-Obeidan S. Comparison between central corneal thickness measurements by oculus pentacam and ultrasonic pachymetry. *Int Ophthalmol*. 2008 Oct;28(5):333-8. Epub 2007 Sep 26.

Barkana Y, Gerber Y, Elbaz U, Schwartz S, Ken-Dror G, Avni I, et al. Central corneal thickness measurement with the Pentacam Scheimpflug system, optical low-coherence reflectometry pachymeter and ultrasound pachymetry. *J Cataract Refract Surg* 2005;31:1729-35.

Behndig A, Mönestam E. Quantification of glistenings in intraocular lenses using Scheimpflug photography. *J Cataract Refract Surg*. 2009 Jan;35(1):14-7.

Belin MW, Khachikian SS. Keratoconus: it is hard to define, but . . . *Am J Ophthalmol* 2007;143:500 –3.

Belin MW and Khachikian SS. An introduction to understanding elevation-based topography: how elevation data are displayed. *Clinical and Experimental Ophthalmology* 2008; 1-15

Belin MW, Khachikian SS, Ambrósio Jr. R, Salomão M. Keratoconus / Ectasia Detection with the Oculus Pentacam: Belin / Ambrósio Enhanced Ectasia Display. *highlights of ophthalmology*, vol. 35, no.6 : 5-12

Bietti G. Ueber familiares Vorkommen von "Retinitis punctata albescens" (verbunden mit "Dystrophia marginaliscristallinea corneae"): Glitzern des Glaskorpers und anderen degenerativen Augenveränderungen. *Klinische Monatsblätter für Augenheilkunde*. 1937;99:737–756.

Binder PS, Lindstrom RL, Stulting RD, et al. Keratoconus and corneal ectasia after LASIK. *J Cataract Refract Surg* 2005;35: 2035–8.

Buehl W, Stojanac D, Sacu S, et al. Comparison of three methods of measuring corneal thickness and anterior chamber depth. *Am J Ophthalmol* 2006;141:7-12.

Cairns G, McGhee CNJ. Orbscan computerized topography: attributes, applications, and limitations. *J Cataract Refract Surg* 2005;31:205–20.

Carpenter J, 1901. Improvements in Enlarging or like Cameras. GB Patent No. 1139. Filed 17 January 1901, and issued 2nd Nov., 1901

Chen D, Lam AKC. Intrasession and intersession repeatability of the Pentacam system on posterior corneal assessment in the normal human eye. *J Cataract Refract Surg* 2007; 33:448–454

Chylack LT, Wolfe JK, Friend J, et al. Validation of methods for the assessment of cataract progression in the Roche European–American Anticataract Trial (REACT). *Ophthalmic Epidemiol* 1995;2:59–75.

Ciolino JB, Khachikian SS, Cortese MJ, Belin MW. Long-term stability of the posterior cornea after Laser In Situ Keratomileusis. *J Cataract Refract Surg*. 2007; 33:1366-70.

Dave T. Current developments in measurement of corneal topography. *Cont Lens Anterior Eye*. 1998;21 Suppl 1:S13-30.

De Sanctis U, Missolungi A, Mutani B, Richiardi L, Grignolo FM, Reproducibility and Repeatability of Central Corneal Thickness Measurement in Keratoconus Using the Rotating Scheimpflug Camera and Ultrasound Pachymetry. *Am J Ophthalmol* 2007;144:712–718.

De Sanctis U, Loiacono C, Richiardi L, Turco D, Mutani B, Grignolo FM. Sensitivity and specificity of posterior corneal elevation measured by Pentacam in discriminating keratoconus/subclinical keratoconus. *Ophthalmology*. 2008 Sep;115(9):1534-9.

Dick HB, Anterior segment analysis for phakic IOL implantation. *Insert to cataract & refractive surgery today*; JAN 2006: 6-7

Dick HB, and Tehrani M. pIOL Simulation for High Res Imaging. *Cataract&Refractive surgery today europe*. January/February 2007

Dubbelman M, Sicam VA, van der Heijde RG. The contribution of the posterior surface to the coma aberration of the human cornea. *J Vis*. 2007 May 30;7(7):10.1-8.

Edmonds CR, Wung S, Pemberton B, Surrett S, Comparison of Anterior Chamber Depth of Normal and Keratoconus Eyes Using Scheimpflug Photography. *Eye & Contact Lens • Volume 35, Number 3, May 2009*. 120-2

Espandar L, Meyer JJ, and Moshirfar M. Phakic intraocular lenses. *Current Opinion in Ophthalmology*. 2008 Jul; 19(4):349–56.

Friedman DS, Gazzard G, Min CB, Broman AT, Quigley H, Tielsch J, Seah S, Foster PJ. Age and sex variation in angle findings among normal Chinese subjects: a comparison of UBM, Scheimpflug, and gonioscopic assessment of the anterior chamber angle. *J Glaucoma* 2008 Jan-Feb;17(1):5-10.

Fujioka M, Nakamura M, Tatsumi Y, Kusuhaara A, Maeda H, Negi A. Comparison of Pentacam Scheimpflug camera with ultrasound pachymetry and noncontact specular microscopy in measuring central corneal thickness. *Curr Eye Res.* 2007 Feb;32(2):89-94.

Fukuda S, Kawana K, Yasuno Y, Oshika T. Anterior Ocular Biometry Using 3-Dimensional Optical Coherence Tomography. *Ophthalmology* 2009;116:882–889

Greste RD. Five in one: an innovation that combines several diagnostic strategies. *Ophthalmology* (special edition). 2004:1-4.

Grewal D, Jain R, Brar GS, Grewal SP. Pentacam tomograms: a novel method for quantification of posterior capsule opacification. *Invest Ophthalmol Vis Sci.* 2008 May;49(5):2004-8.

Grewal DS, Brar GS, Grewal SP. Correlation of Nuclear Cataract Lens Density using Scheimpflug Images with Lens Opacities Classification System III and Visual Function. *Ophthalmology* 2009 Jun 3.

Grewal SP, Jain R, Gupta R, Grewal D. Role of scheimpflug imaging in traumatic intralenticular foreign body. *Am J Ophthalmol.* 2006 Oct;142(4):675-6.

Ha BJ, Kim SW, Kim SW, Kim EK, Kim T. Pentacam and Orbscan II Measurements of Posterior Corneal Elevation Before and After Photorefractive Keratectomy. [*J Refract Surg.* 2009;25:290-295.

Ho T, Cheng AC, Rao SK, Lau S, Leung CK, Lam DS. Central corneal thickness measurements using Orbscan II, Visante, ultrasound, and Pentacam pachymetry after laser in situ keratomileusis for myopia. *J Cataract Refract Surg* 2007;33:1177–1182.

Hockwin O, Dragomirescu V, Laser H, Wegener A. Scheimpflug photography of the anterior eye segment. Principle, instrumentation and application to clinical and experimental ophthalmology. *J Toxicol Cutaneous Ocul Toxicol* 1987; 6: 251–71.

Holladay JT, Belin MW, Chayet AS, Maus M, Vinciguerra P. Next generation technology for cataract&refractive surgeon. *Cataract and refractive surgery today.* 2005;january(suppl): 1-11

Holladay JT. Holladay Report on the Pentacam. *New Advances and Technology with Pentacam;* 2008: 10-15

Hu CY, Jian JH, Cheng YP, Hsu HK. Analysis of crystalline lens position. *J Cataract Refrac Surg* 2006;32:599–603.

Jain R, Grewal D, Gupta R, Grewal SP. Scheimpflug imaging in late Capsular Bag Distention syndrome after phacoemulsification. *Am J Ophthalmol*. 2006 Dec;142(6):1083-5.

Jain R, Dilraj G, Grewal SP. Repeatability of corneal parameters with Pentacam after laser in situ keratomileusis. *Indian J Ophthalmol*. 2007 Sep-Oct;55(5):341-7.

Kawamorita T, Uozato H, Kamiya K, Bax L, Tsutsui K, Aizawa D, Shimizu K. Repeatability, reproducibility, and agreement characteristics of rotating Scheimpflug photography and scanning-slit corneal topography for corneal power measurement. *J Cataract Refract Surg*. 2009 Jan;35(1):127-133.

Kim SW, Byun YJ, Kim EK, Kim TI. Central corneal thickness measurements in unoperated eyes and eyes after PRK for myopia using Pentacam, Orbscan II, and ultrasonic pachymetry. *J Refract Surg*. 2007 Nov;23(9):888-94

Kitagawa, K; Sakamoto, Y; Sasaki, K; Hanaki, H. Evaluation of transparency and barrier function of the cornea by Scheimpflug images. *Ophthalmic research* vol. 28 Suppl 2, (1996): 72-7

Konstantopoulos A, Hossain P, Anderson DF. Recent advances in ophthalmic anterior segment imaging: a new era for ophthalmic diagnosis?. *Br J Ophthalmol* 2007;91:551–557.

Lackner B, Schmidinger G, Skorpik C. Validity and repeatability of anterior chamber depth measurements with pentacam and orbscan. *Optom Vis Sci* 2005;82:858-61.

Lam AK, Chan R, Woo GC, et al: Intra-observer and interobserver repeatability of anterior eye segment analysis system (EAS-1000) in anterior chamber configuration. *Ophthalmic Physiol Opt* 22:552--9, 2002

Lam AK, Chen D. Pentacam pachometry: comparison with non-contact specular microscopy on the central cornea and inter-session repeatability on the peripheral cornea. *Clin Exp Optom*. 2007 Mar;90(2):108-14.

Latkany RA, Chokshi AR, Speaker MG, et al. Intraocular lens calculations after refractive surgery. *J Cataract Refract Surg* 2005; 31:562–570

Li X, Rabinowitz YS, Rasheed K, Yang H. Longitudinal study of the normal eyes in unilateral keratoconus patients. *Ophthalmology* 2004;111:440–6.

Matsuda J, Hieda O, Kinoshita S. Comparison of central corneal thickness measurements by Orbscan II and Pentacam after corneal refractive surgery. *Jpn J Ophthalmol*. 2008 Jul-Aug;52(4):245-9. Epub 2008 Sep 5.

Menassa N, Kaufmann C, Goggin M, Job OM, Bachmann LM, Thiel MA. Comparison and reproducibility of corneal thickness and curvature readings obtained by the GALILEI™ and the Orbscan II analysis systems. *J Cataract Refract Surg*. 2008 Oct; 34(10):1742-7.

Merklinger, Harold M. 1996. *Focusing the View Camera*. Bedford, Nova Scotia: Seaboard Printing Limited. ISBN 0-9695025-2-4.

Merklinger, Harold M. 2002. *The INs and OUTs of focus*. Bedford, Nova Scotia: Seaboard Printing Limited. ISBN 0-9695025-0-8.

Mirjam S. Reuland, MD; Andreas J. Reuland, MD; Yutaro Nishi, MD; Gerd U. Auffarth, MD. Corneal Radii and Anterior Chamber Depth Measurements Using the IOLMaster Versus the Pentacam. *J Refract Surg*. 2007;23:368-373.

Nakagawa T, Maeda N, Kosaki R, Hori Y, Ioue T, Saika M, Mihashi T, Fujikado T, Tano Y. Higher-Order Aberrations due to the Posterior Corneal Surface in Patients with Keratoconus. *Invest Ophthalmol Vis Sci*. 2008 Nov 21.

O'Donnell C, Maldonado-Codina C. Agreement and repeatability of central thickness measurement in normal corneas using ultrasound pachymetry and the OCULUS Pentacam. *Cornea* 2005;24:920-4.

Pei X, Bao Y, Chen Y, Li X. Correlation of lens density measured using the Pentacam Scheimpflug system with the Lens Opacities Classification System III grading score and visual acuity in age-related nuclear cataract. *Br J Ophthalmol*. 2008 Nov;92(11):1471-5.

Piñero DP, Gonza'lez CS, Alio' JL. Intraobserver and interobserver repeatability of curvature and aberrometric measurements of the posterior corneal surface in normal eyes using Scheimpflug photography. *J Cataract Refract Surg*. 2009 Jan;35(1):113-120.

Rabinowitz YS, Rasheed K. KISA% index: a quantitative videokeratography algorithm embodying minimal topographic criteria for diagnosing keratoconus. *J Cataract Refract Surg* 1999;25:1327–35.

Rabsilber TM, Khoramnia R, Auffarth GU. Anterior chamber measurements using Pentacam rotating Scheimpflug camera. *J Cataract Refract Surg*. 2006 Mar;32(3):456-9.

Reddy RA, Pande MV, Finn P, et al. Comparative estimation of anterior chamber depth by ultrasonography, Orbscan II, and IOL Master. *J Cataract Refract Surg* 2004;30:1268–1271

- Salouti R, Nowroozadeh MH, Zamani M, Fard AH, Niknam S. Comparison of anterior and posterior elevation map measurements between 2 Scheimpflug imaging systems. *J Cataract Refract Surg* 2009; 35:856–862
- Sasaki K, Sakamoto Y, Shibata T, Emori Y. The multipurpose camera: a new anterior eye segment analysis system. *Ophthalmic Res* 1990; 22 (Suppl. 1): 3–8.
- Sasaki H, Kawakami Y, Ono M et al. Localization of cortical cataract in subjects of diverse races and latitude. *Invest Ophthalmol Vis Sci* 2003;44:4210–4214.
- Savant V, Chavan R, Pushpoth S, Ilango B. Comparability and Intra-/Interobserver Reliability of Anterior Chamber Depth Measurements With the Pentacam and IOLMaster. *J Refract Surg*. 2008;24:615-618.
- Savini G, Barboni P, Profazio V, Zanini M, Hoffer KJ. Corneal power measurements with the Pentacam Scheimpflug camera after myopic excimer laser surgery. *J Cataract Refract Surg* 2008; 34:809–813
- Scheimpflug T, 1904. Improved Method and Apparatus for the Systematic Alteration or Distortion of Plane Pictures and Images by Means of Lenses and Mirrors for Photography and for other purposes. GB Patent No. 1196. Filed 16 January 1904, and issued 12 May 1904
- Schumacher S, Fromm M, Oberheide U, Gerten G, Wegener A, Lubatschowski H. In vivo application and imaging of intralenticular femtosecond laser pulses for the restoration of accommodation. *J Refract Surg*. 2008 Nov;24(9):991-5.
- Shankar H, Taranath D, Santhirathelagan CT, Pesudovs K. Repeatability of corneal first-surface wavefront aberrations measured with Pentacam corneal topography. *J Cataract Refract Surg* 2008; 34:727–734
- Swartz T, Marten L and Wang M. Measuring the cornea: the latest developments in corneal topography. *Current Opinion in Ophthalmology* 2007, 18:325–333
- Turner SJ, Lee EJK, Hu V, Hollick EJ. Scheimpflug imaging to determine intraocular lens power in vivo. *J Cataract Refract Surg* 2007; 33:1041–1044.
- Uttine CA, Altin F, Cakir H and Perente I. Comparison of anterior chamber depth measurements taken with the Pentacam, Orbscan IIz and IOLMaster in myopic and emmetropic eyes. *Acta Ophthalmol*. 2009 Jun;87(4):386-91.
- Walker RN, Khachikian SS, Belin MW. Scheimpflug photographic diagnosis of pellucid marginal degeneration. *Cornea*. 2008 Sep;27(8):963-6
- Welch RB. Bietti's tapetoretinal degeneration with marginal corneal dystrophy crystalline retinopathy. *Trans Am Ophthalmol Soc*. 1977;75:164–179.

